

CW medical GbR
Buldernweg 57
DE-48163 Münster

Hamburg, 23 June 2025

Expert opinion

Activity of **Virosan Flächen-Desinfektionsmittel ohne Alkohol** in the
quantitative suspension test according to DIN EN 17126:2019-02

The Disinfectant **Virosan Flächen-Desinfektionsmittel ohne Alkohol** was tested and evaluated according to DIN EN 17126:2019-02 "Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of sporicidal activity in the medical area – Test method and requirements (Phase 2/Step 1); German version EN 17126:2019".

According to the test report no. L25/000287.1 dated 23/06/2025 of Dr. Brill + Partner GmbH the preparation showed sporicidal activity against *Clostridioides difficile*, *Bacillus spizizenii (subtilis)* and *Bacillus cereus* under clean conditions (0.3 g/L bovine albumin) at 20 °C.

Virosan Flächen-Desinfektionsmittel ohne Alkohol complies with the requirements of DIN EN 17126:2019-02 with the following concentration-time relationship:

Claim	Organic load	Concentration	Contact time
Sporicidal (<i>B. spizizenii</i> and <i>B. cereus</i>)	clean conditions	100 %	15 minutes



Dipl.-Biol. Dr. rer. nat. Florian H. H. Brill